

DECLERATION OF CONFORMITY

In accordance with EN ISO 17050 - 1:2004

We Evacusafe UK Limited

Of 9, Devonshire Business Park

Knights Park Road

Basingstoke Hampshire RG21 6XN

In accordance with the following Directive:

93/42/EEC The Medical Devices Directive and its amending directives

Hereby declare that:

Equipment Evacuation Equipment

Model Nos Excel, Standard, 2W Transit Chair, 4W Transit Chair

Home Emergency Chair (Tri-Wheel), Basic Slider, Premium Adjustable Slider

Classification Class 1 Medical Devices

Is in conformity with the applicable requirements of the following documents:

REF No	TITLE	EDITION/DATE
EN 980	Graphical symbols for use in the labelling of medical devices.	2008
EN 1041	Information supplied by the manufacturer with medical devices.	1998
EN 14971	Medical devices – Application of risk management to medical devices.	2007

I hereby declare that the equipment named above has been designed to comply with the relevant sections of the above referenced specifications. The units comply with all applicable Essential Requirements of the Directives.

Signed by:

Name: Paul Mitchell

Position: Director

On: 07/11/14

